Calibration Certificate

Certificate No. TW241011_04

		Desci	ription		
		Unit Name	Vaccine Refrigerator		
Calibration Subject	Temperature Sensor	Model No.	MPR-1411R	Manufacture	PANASONIC
		Serial No.	14060339		
Date of C	Calibration		2024-1	0-08	

		Calibraion Method	
Utilized Equipment	Model	Serial No	Manufacture
Utilized Equipment	GR200-12NO	10DPM0203	HANYOUNG NUX

	Calibrati	on Result		
Tomporatura Cat Daint	Measur	ed Value	Talayanas	Danilt
Temperature Set Point	Refrigerator	Recorder	Tolerance	Result
5.0℃	5.0℃	4.9℃	±2 ℃	Pass
AFFIRMATION		Performed by :	TWINMEDITECH	

확인 의뢰자 (Client) 승인자 (Approved) Confirmation 해운대백병원 Name & Signature: 전진환
--

본 장비의 Calibration이 아래 일자로 실행 되었음을 증명합니다.

2024년 10월 11일



