

Calibration Certificate

Certificate No. TW241011_03

Description					
Calibration Subject	Temperature Sensor	Unit Name	Vaccine Refrigerator	Manufacture	(주)BRM
		Model No.	NPH-1407L		
		Serial No.	H14072201-002		
Date of Calibration		2024-10-08			

Calibraion Method			
Utilized Equipment	Model	Serial No	Manufacture
	GR200-12NO	10DPM0203	HANYOUNG NUX

Calibration Result				
Temperature Set Point	Measured Value		Tolerance	Result
	Refrigerator	Recorder		
5.0°C	4.0°C	3.8°C	±2 °C	Pass
AFFIRMATION	Performed by : TWINMEDITECH			

확인 Confirmation	의뢰자 (Client) 해운대백병원	승인자 (Approved) Name & Signature: 전진환
--------------------	------------------------	---

본 장비의 Calibration이 아래 일자로 실행 되었음을 증명합니다.

2024년 10월 11일



트윈메디텍

